

Calibration Gases used in Flue Gas Testing of Unit 2

<u>Location</u>	<u>Gas</u>	<u>use</u>	<u>Concentration</u>
Eleventh Floor Test Station	Nitrogen	zero gas	99.99% pure
Eleventh Floor Test Station	CO	Carbon Monoxide Calibration	1000 ppm
Eleventh Floor Test Station	CO ₂	Carbon Dioxide Calibration	18.88%
Eleventh Floor Test Station	O ₂	Oxygen Span Gas Calibration	8.01%
Eleventh Floor Test Station	O ₂	Oxygen Zero Gas Calibration	0.40%
Stack	CO	Carbon Monoxide Calibration	997 ppm
Stack	O ₂	Oxygen Span Gas Calibration	8.00%
Stack	O ₂	Oxygen Zero Gas Calibration	0.40%
4th Floor West Side Test Station	Nitrogen	zero gas	99.99% pure
5th Floor West Side Test Station	CO ₂	Carbon Dioxide Calibration	18.88%
6th Floor West Side Test Station	O ₂	Oxygen Span Gas Calibration	8.02%
4th Floor East Side Test Station	Nitrogen	zero gas	99.99% pure
5th Floor East Side Test Station	CO ₂	Carbon Dioxide Calibration	18.92%
6th Floor East Side Test Station	O ₂	Oxygen Span Gas Calibration	8.02%

INVENTORY CONTROL		STORES FILE INQ		IC0461	
Stock Number: 38381				Page 1 of 9	
1	Noun	NITROGEN	Size		
Qualifier:					
	Primary LOC	GSB GSBWHS		New	Cap
2	Row/Bin	LDBR01N	Binned:	2	0
	Buyer:	06 LESLIE LOVELL	Rcvd/Unbinned:		0
3	Unit of Issue:	BT	Total on hand:		2
4	Avg Unit Cost:	101.20000	Reserved:		0
5	Reb Avg Unit C:		Available:		2
	Cap Spare Val:		On Reqs:		0
6	Class (1-10):	51 LIMITED ACCESS(PUR)	On Order:		6
7	Item EMB:	M	Qty On Loan:		0
8	Serial Flag:	N	Qty In Shop:		0
9	Obso/Surplus:		13 Reb/Fab Unbin:		
10	MFG Part No:	BOTTLE	Net In System:		8
11	Manufacturer	US WELDING	New Inv Val:		202.40
12	Description:	NITROGEN, 99.99% PURE, CYLINDER SIZE 304, FT3, T SIZE, US WELDING BOTTLE LIMITED TO LAB			

WHICH ONE (E=EXIT, P#=PAGE, M=MSDS, W=WHERE USED) ?

D54600 10:17 29 APR 2004 Intermountain Power Service Corp. USER.LIVE.DATA

1414 Floor2

PRAXAIR

Praxair
5700 South Alameda Street
Los Angeles, CA 90058
Telephone: (323) 585-2154
Facsimile: (714) 542-6689

CERTIFICATE OF ANALYSIS

CUSTOMER IPSC

DATE 03/22/04

P.O NUMBER

REF. NUMBER 17650100

REQUESTED COMPOSITION

GAS	CONCENTRATION
CARBON MONOXIDE	1000 ppm
NITROGEN	BALANCE
ANALYTICAL ACCURACY	±1 %

ANALYTICAL METHOD

INSTRUMENT	ANALYTICAL PRINCIPLE
HORIBA, VIA-510, S/N 576876015	NDIR

Values not valid below 150 psig.
This certified mix is N.I.S.T traceable to
SRM#1681b for carbon monoxide.

THIS CYLINDER NO.		CERTIFIED CONCENTRATION	
CYLINDER PRESSURE	2000 PSIG	CARBON MONOXIDE	1000 ppm
EXPIRATION DATE	03/22/07	NITROGEN	BALANCE
CLASSIFICATION	PRIMARY MASTER	ANALYTICAL ACCURACY	±1 %
BATCH NUMBER	N/A		
LOT NUMBER	109407903		
PART NUMBER	EV NICO1000P-AS		
CYLINDER SIZE	AS CGA 350		
	143 CFT		

ANALYZED BY

JACK FU

CERTIFIED BY

TY TRIPLETT

IMPORTANT

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any particular purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc., arising out of the use of the information contained herein exceed the fee established for providing such information.

IP7_035306



Praxair
5700 South Alameda Street
Los Angeles, CA 90058
Telephone: (323) 585-2154
Facsimile: (714) 542-6689

CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS

CUSTOMER IPSC

P.O NUMBER

REFERENCE STANDARD

COMPONENT	NIST SRM NO.	CYLINDER NO.	CONCENTRATION
NITRIC OXIDE GMIS	vsSRM#1687b	CC 108618	1005 ppm
SULFUR DIOXIDE	vs.SRM1662a	CC 16163	1012 ppm
CARBON DIOXIDE GMIS	vs.SRM#2745	8747556	17.99 %

ANALYZER READINGS

R=REFERENCE STANDARD

Z=ZERO GAS

C=GAS CANDIDATE

1. COMPONENT	NITRIC OXIDE	GMIS	ANALYZER MAKE-MODEL-S/N	HORIBA, VIA-510, 57042305
ANALYTICAL PRINCIPLE	NDIR			LAST CALIBRATION DATE 09/02/03
FIRST ANALYSIS DATE	08/26/03			SECOND ANALYSIS DATE 09/03/03
Z 0	R 1001	C 881.3	CONC. 885	Z 0 R 1005 C 876.3 CONC. 876
R 1001	Z 0	C 879.3	CONC. 883	R 1005 Z 0 C 876.3 CONC. 876
Z 0	C 881.3	R 1001	CONC. 885	Z 0 C 876.3 R 1005 CONC. 876
U/M ppm		MEAN TEST ASSAY	884	U/M ppm MEAN TEST ASSAY 876
2. COMPONENT	SULFUR DIOXIDE		ANALYZER MAKE-MODEL-S/N	Siemens Ultramat 5E S/N C1-009
ANALYTICAL PRINCIPLE	NDIR			LAST CALIBRATION DATE 08/11/03
FIRST ANALYSIS DATE	08/26/03			SECOND ANALYSIS DATE 09/03/03
Z 0	R 1012	C 868	CONC. 868	Z 0 R 1012 C 868 CONC. 868
R 1012	Z 0	C 868	CONC. 868	R 1012 Z 0 C 870 CONC. 870
Z 0	C 868	R 1012	CONC. 868	Z 0 C 870 R 1012 CONC. 870
U/M ppm		MEAN TEST ASSAY	868	U/M ppm MEAN TEST ASSAY 869
3. COMPONENT	CARBON DIOXIDE	GMIS	ANALYZER MAKE-MODEL-S/N	HORIBA/VIA-510/SN#02807014
ANALYTICAL PRINCIPLE	NDIR			LAST CALIBRATION DATE 08/11/03
FIRST ANALYSIS DATE	08/26/03			SECOND ANALYSIS DATE
Z 0	R 17.98	C 18.88	CONC. 18.89	Z R C CONC.
R 17.98	Z 0	C 18.86	CONC. 18.87	R Z C CONC.
Z 0	C 18.86	R 17.98	CONC. 18.87	Z C R CONC.
U/M %		MEAN TEST ASSAY	18.88	U/M % MEAN TEST ASSAY

THIS CYLINDER NO. SA 16967

HAS BEEN CERTIFIED ACCORDING TO SECTION

OF TRACEABILITY PROTOCOL NO.

PROCEDURE G1

CERTIFIED ACCURACY ± 1 % NIST TRACEABLE

CYLINDER PRESSURE 2000 PSIG

CERTIFICATION DATE 09/03/03

EXPIRATION DATE 09/03/05 TERM 24 MONTHS

CERTIFIED CONCENTRATION

NITRIC OXIDE	880 ppm
SULFUR DIOXIDE	868 ppm
CARBON DIOXIDE	18.88 %
NITROGEN	BALANCE

NOx=880 ppm, values for reference only. Values not valid
below 150 psig. NO conc. is corrected for CO2 interferenc

ANALYZED BY

JOSEPH CHARLES

CERTIFIED BY

JACK FE

IMPORTANT

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any particular purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc., arising out of the use of the information contained herein exceed the fee established for providing such information.

IP7_035307



Praxair
5700 South Alameda Street
Los Angeles, CA 90058
Telephone: (323) 585-2154
Facsimile: (714) 542-6689

CERTIFICATE OF ANALYSIS

CUSTOMER IPSC

DATE 01/26/04

P.O NUMBER

REF. NUMBER 61290100

REQUESTED COMPOSITION

GAS	CONCENTRATION
OXYGEN	8 %
NITROGEN	BALANCE
ANALYTICAL ACCURACY	±1 %

ANALYTICAL METHOD

INSTRUMENT	ANALYTICAL PRINCIPLE
Siemens Oxymat 5E S/N A12-839	Paramagnetic

Values not valid below 150 psig.

This certified mix is N.I.S.T traceable to SRM#2659a for oxygen.

THIS CYLINDER NO.	CC 80585
CYLINDER PRESSURE	2000 PSIG
EXPIRATION DATE	01/23/07
CLASSIFICATION	PRIMARY MASTER
BATCH NUMBER	N/A
LOT NUMBER	109402012
PART NUMBER	EV NIOX8N-AS
CYLINDER SIZE	AS CGA 590 144 CFT

CERTIFIED CONCENTRATION

OXYGEN	8.01 %
NITROGEN	BALANCE
ANALYTICAL ACCURACY	±1 %

ANALYZED BY

JACK FU

CERTIFIED BY

TY TRIPLETT

IMPORTANT

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any particular purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc., arising out of the use of the information contained herein exceed the fee established for providing such information.

IP7_035308

Praxair
5700 South Alameda Street
Los Angeles, CA 90058
Telephone: (323) 585-2154
Facsimile: (714) 542-6689

CERTIFICATE OF ANALYSIS

CUSTOMER IPSC

DATE 06/20/03

P.O NUMBER

REF. NUMBER 69661200

REQUESTED COMPOSITION

GAS	CONCENTRATION
OXYGEN	0.4 %
NITROGEN	BALANCE
ANALYTICAL ACCURACY	±1 %

ANALYTICAL METHOD

INSTRUMENT	ANALYTICAL PRINCIPLE
METTLER ID5, S/N:1865166	GRAVIMETRIC

Values not valid below 150 psig.

THIS CYLINDER NO.	CC 131744
CYLINDER PRESSURE	2000 PSIG
EXPIRATION DATE	12/20/03
CLASSIFICATION	PRIMARY STANDARD
BATCH NUMBER	N/A
LOT NUMBER	109317010
PART NUMBER	NI OX.4P-AS
CYLINDER SIZE	AS CGA 580 143 CFT

CERTIFIED CONCENTRATION

OXYGEN	0.400 %
NITROGEN	BALANCE
ANALYTICAL ACCURACY	±1 %

ANALYZED BY

JACK FU

CERTIFIED BY

EDGAR CUEVAS

IMPORTANT

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any particular purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc., arising out of the use of the information contained herein exceed the fee established for providing such information.

IP7_035309

Stack



Praxair
5700 South Alameda Street
Los Angeles, CA 90058
Telephone: (323) 585-2154
Facsimile: (714) 542-6689

CERTIFICATE OF ANALYSIS

CUSTOMER IPSC

DATE 03/22/04

P.O NUMBER

REF. NUMBER 17650100

REQUESTED COMPOSITION

GAS	CONCENTRATION
CARBON MONOXIDE	1000 ppm
NITROGEN	BALANCE
ANALYTICAL ACCURACY	±1 %

ANALYTICAL METHOD

INSTRUMENT

HORIBA, VIA-510, S/N 576876015

ANALYTICAL PRINCIPLE

NDIR

Values not valid below 150 psig.

This certified mix is N.I.S.T traceable to
SRM#1681b for carbon monoxide.

THIS CYLINDER NO.	CC 163418
CYLINDER PRESSURE	2000 PSIG
EXPIRATION DATE	03/22/07
CLASSIFICATION	PRIMARY MASTER
BATCH NUMBER	N/A
LOT NUMBER	109407901
PART NUMBER	EV NICO1000P-AS
CYLINDER SIZE	AS CGA 350 143 CFT

CERTIFIED CONCENTRATION

CARBON MONOXIDE	997 ppm
NITROGEN	BALANCE
ANALYTICAL ACCURACY	±1 %

ANALYZED BY

JACK FU

CERTIFIED BY

TY TRIPLETT

IMPORTANT

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any particular purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc., arising out of the use of the information contained herein exceed the fee established for providing such information.

IP7_035310

Stack

PRAXAIR

Praxair
5700 South Alameda Street
Los Angeles, CA 90058
Telephone: (323) 585-2154
Facsimile: (714) 542-6689

CERTIFICATE OF ANALYSIS

CUSTOMER IPSC

DATE 03/03/04

P.O NUMBER

REF. NUMBER 96495600

REQUESTED COMPOSITION

GAS	CONCENTRATION
OXYGEN	8 %
NITROGEN	BALANCE
ANALYTICAL ACCURACY	±1 %

ANALYTICAL METHOD

INSTRUMENT	ANALYTICAL PRINCIPLE
Siemens Oxymat 5E S/N A12-839	Paramagnetic

Values not valid below 150 psig.
This certified mix is N.I.S.T traceable to
SRM#2658a for oxygen.

THIS CYLINDER NO.	CC 145491
CYLINDER PRESSURE	2000 PSIG
EXPIRATION DATE	03/03/07
CLASSIFICATION	PRIMARY MASTER
BATCH NUMBER	N/A
LOT NUMBER	109406104
PART NUMBER	EV NIOX8N-AS
CYLINDER SIZE	AS CGA 590 144 CFT


CERTIFIED CONCENTRATION

OXYGEN	7.99 %
NITROGEN	BALANCE
ANALYTICAL ACCURACY	±1 %

ANALYZED BY


JACK FU

CERTIFIED BY


TY TRIPLETT

IMPORTANT

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any particular purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc., arising out of the use of the information contained herein exceed the fee established for providing such information.

IP7_035311

INVENTORY CONTROL

STORES FILE INQ

IC0461

Stock Number: 49189

Page 1 of 9

1 Noun	GAS	Type	Concentrat	Cyl. Size	
Qualifier:	CALIBRATION	O2	0.4%	150 SCF	
Primary LOC	GSB GSBWHS			New	Cap Reb
2 Row/Bin	LD14G		Binned:	1	0 0
Buyer:	05 NANCY BENNETT		Rcvd/Unbinned:		0
3 Unit of Issue:	EA		Total on hand:		1
4 Avg Unit Cost:	138.18660		Reserved:		0
5 Reb Avg Unit C:			Available:		1
Cap Spare Val:			On Reqs:		0
6 Class (1-10):	51 LIMITED ACCESS(PUR)		On Order:		2
7 Item EMB:	B		Qty On Loan:		0
8 Serial Flag:	N		Qty In Shop:		0
9 Obso/Surplus:		13	Reb/Fab Unbinn:		
10 MFG Part No:	150-O2-0.4		Net In System:		3
11 Manufacturer	SCOTT MARRIN		New Inv Val:		138.19
12 Description:	GAS, CALIBRATION, O2, 0.4%, 150 SCF, (limited access to I&C technicians only)				
	SCOTT MARRIN / 150-O2-0.4				

WHICH ONE (E=EXIT, P#=PAGE, M=MSDS, W=WHERE USED) ?

D54600 10:18 29 APR 2004 Intermountain Power Service Corp. USER.LIVE.DATA

IP7_035312



Praxair
5700 South Alameda Street
Los Angeles, CA 90058
Telephone: (323) 585-2154
Facsimile: (714) 542-6689

CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS

CUSTOMER IPSC

P.O NUMBER

REFERENCE STANDARD

COMPONENT	NIST SRM NO.	CYLINDER NO.	CONCENTRATION
NITRIC OXIDE GMIS	vsSRM#1687b	CC 108618	1005 ppm
SULFUR DIOXIDE	vs.SRM1662a	CC 16163	1012 ppm
CARBON DIOXIDE GMIS	vs.SRM#2745	8747556	17.99 %

ANALYZER READINGS

R=REFERENCE STANDARD

Z=ZERO GAS

C=GAS CANDIDATE

1. COMPONENT	NITRIC OXIDE	GMIS	ANALYZER MAKE-MODEL-S/N	HORIBA, VIA-510, 57042305			
ANALYTICAL PRINCIPLE	NDIR		LAST CALIBRATION DATE	09/02/03			
FIRST ANALYSIS DATE	08/26/03		SECOND ANALYSIS DATE	09/03/03			
Z 0	R 1001	C 876.3	CONC. 880	Z 0	R 1005	C 878.3	CONC. 878
R 1001	Z 0	C 873.3	CONC. 877	R 1005	Z 0	C 879.3	CONC. 879
Z 0	C 874.3	R 1001	CONC. 878	Z 0	C 879.3	R 1005	CONC. 879
U/M	ppm	MEAN TEST ASSAY	878	U/M	ppm	MEAN TEST ASSAY	879
2. COMPONENT	SULFUR DIOXIDE		ANALYZER MAKE-MODEL-S/N	Siemens Ultramat SE S/N C1-009			
ANALYTICAL PRINCIPLE	NDIR		LAST CALIBRATION DATE	08/11/03			
FIRST ANALYSIS DATE	08/26/03		SECOND ANALYSIS DATE	09/03/03			
Z 0	R 1012	C 866	CONC. 866	Z 0	R 1012	C 864	CONC. 864
R 1012	Z 0	C 866	CONC. 866	R 1012	Z 0	C 864	CONC. 864
Z 0	C 866	R 1012	CONC. 866	Z 0	C 864	R 1012	CONC. 864
U/M	ppm	MEAN TEST ASSAY	866	U/M	ppm	MEAN TEST ASSAY	864
3. COMPONENT	CARBON DIOXIDE	GMIS	ANALYZER MAKE-MODEL-S/N	HORIBA/VIA-510/SN#02807014			
ANALYTICAL PRINCIPLE	NDIR		LAST CALIBRATION DATE	08/11/03			
FIRST ANALYSIS DATE	08/26/03		SECOND ANALYSIS DATE				
Z 0	R 17.98	C 18.84	CONC. 18.85	Z	R	C	CONC.
R 17.98	Z 0	C 18.92	CONC. 18.93	R	Z	C	CONC.
Z 0	C 18.86	R 17.98	CONC. 18.87	Z	C	R	CONC.
U/M	%	MEAN TEST ASSAY	18.88	U/M	%	MEAN TEST ASSAY	

THIS CYLINDER NO. CC 19125
HAS BEEN CERTIFIED ACCORDING TO SECTION EPA-600/R97/121
OF TRACEABILITY PROTOCOL NO. Rev. 9/97
PROCEDURE G1
CERTIFIED ACCURACY ± 1 % NIST TRACEABLE
CYLINDER PRESSURE 2000 PSIG
CERTIFICATION DATE 09/03/03
EXPIRATION DATE 09/03/05 TERM 24 MONTHS

CERTIFIED CONCENTRATION
NITRIC OXIDE 878 ppm
SULFUR DIOXIDE 865 ppm
CARBON DIOXIDE 18.88 %
NITROGEN BALANCE

NOx=878 ppm, values for reference only. Values not valid
below 150 psig. NO conc. is corrected for CO2 interferenc

ANALYZED BY

JOSEPH CHARLES

CERTIFIED BY

JACK FU

IMPORTANT

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any particular purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc., arising out of the use of the information contained herein exceed the fee established for providing such information.

IP7_035313

West Side 444



Praxair
5700 South Alameda Street
Los Angeles, CA 90058
Telephone: (323) 585-2154
Facsimile: (714) 542-6689

CERTIFICATE OF ANALYSIS

CUSTOMER IPSC

DATE 01/26/04

P.O NUMBER

REF. NUMBER 61290100

REQUESTED COMPOSITION

GAS	CONCENTRATION
OXYGEN	8 %
NITROGEN	BALANCE
ANALYTICAL ACCURACY	±1 %

ANALYTICAL METHOD

INSTRUMENT
Siemens Oxymat 5E S/N A12-839

ANALYTICAL PRINCIPLE
Paramagnetic

Values not valid below 150 psig.

This certified mix is N.I.S.T traceable to SRM#2659a for oxygen.

THIS CYLINDER NO.	CC 108622
CYLINDER PRESSURE	2000 PSIG
EXPIRATION DATE	01/23/07
CLASSIFICATION	PRIMARY MASTER
BATCH NUMBER	N/A
LOT NUMBER	109402010
PART NUMBER	EV NIOX8N-AS
CYLINDER SIZE	AS CGA 590 144 CFT

CERTIFIED CONCENTRATION

OXYGEN	8.02 %
NITROGEN	BALANCE
ANALYTICAL ACCURACY	±1 %

ANALYZED BY

JACK ED

CERTIFIED BY

TY TRIPLETT

IMPORTANT

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any particular purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc., arising out of the use of the information contained herein exceed the fee established for providing such information.

IP7_035314



Praxair
5700 South Alameda Street
Los Angeles, CA 90058
Telephone: (323) 585-2154
Facsimile: (714) 542-6689

CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS

CUSTOMER IPSC

P.O NUMBER

REFERENCE STANDARD

COMPONENT	NIST SRM NO.	CYLINDER NO.	CONCENTRATION
NITRIC OXIDE GMIS	vsSRM#1687b	CC 108618	1005 ppm
SULFUR DIOXIDE	vs.SRM1662a	CC 16163	1012 ppm
CARBON DIOXIDE GMIS	vs.SRM#2745	8747556	17.99 %

ANALYZER READINGS

R=REFERENCE STANDARD

Z=ZERO GAS

C=GAS CANDIDATE

1. COMPONENT	NITRIC OXIDE	GMIS	ANALYZER MAKE	MODEL-S/N	HORIBA, VIA-510, 57042305				
ANALYTICAL PRINCIPLE	NDIR				LAST CALIBRATION DATE	09/02/03			
FIRST ANALYSIS DATE	08/26/03				SECOND ANALYSIS DATE	09/03/03			
Z 0	R 1001	C 875.3	CONC.	879	Z 0	R 1005	C 875.3	CONC.	875
R 1001	Z 0	C 875.3	CONC.	879	R 1005	Z 0	C 876.3	CONC.	876
Z 0	C 875.3	R 1001	CONC.	879	Z 0	C 876.3	R 1005	CONC.	876
U/M	ppm	MEAN TEST ASSAY	879	U/M	ppm	MEAN TEST ASSAY	876		
2. COMPONENT	SULFUR DIOXIDE		ANALYZER MAKE	MODEL-S/N	Siemens Ultramat SE S/N C1-009				
ANALYTICAL PRINCIPLE	NDIR				LAST CALIBRATION DATE	08/11/03			
FIRST ANALYSIS DATE	08/26/03				SECOND ANALYSIS DATE	09/03/03			
Z 0	R 1012	C 866	CONC.	866	Z 0	R 1012	C 866	CONC.	866
R 1012	Z 0	C 866	CONC.	866	R 1012	Z 0	C 868	CONC.	868
Z 0	C 866	R 1012	CONC.	866	Z 0	C 868	R 1012	CONC.	868
U/M	ppm	MEAN TEST ASSAY	867	U/M	ppm	MEAN TEST ASSAY	868		
3. COMPONENT	CARBON DIOXIDE	GMIS	ANALYZER MAKE	MODEL-S/N	HORIBA/VIA-510/SN#02807014				
ANALYTICAL PRINCIPLE	NDIR				LAST CALIBRATION DATE	08/11/03			
FIRST ANALYSIS DATE	08/26/03				SECOND ANALYSIS DATE				
Z 0	R 17.98	C 18.88	CONC.	18.89	Z	R	C	CONC.	
R 17.98	Z 0	C 18.94	CONC.	18.95	R	Z	C	CONC.	
Z 0	C 18.90	R 17.98	CONC.	18.91	Z	C	R	CONC.	
U/M	%	MEAN TEST ASSAY	18.92	U/M	%	MEAN TEST ASSAY			

THIS CYLINDER NO. CC 102202
HAS BEEN CERTIFIED ACCORDING TO SECTION EPA-600/R97/121
OF TRACEABILITY PROTOCOL NO. Rev. 9/97
PROCEDURE G1
CERTIFIED ACCURACY ± 1 % NIST TRACEABLE
CYLINDER PRESSURE 2000 PSIG
CERTIFICATION DATE 09/03/03
EXPIRATION DATE 09/03/05 TERM 24 MONTHS

CERTIFIED CONCENTRATION

NITRIC OXIDE	878 ppm
SULFUR DIOXIDE	868 ppm
CARBON DIOXIDE	18.92 %
NITROGEN	BALANCE

NOx=878 ppm, values for reference only. Values not valid
below 150 psig. NO conc. is corrected for CO2 interferenc

ANALYZED BY

JOSEPH CHARLES

CERTIFIED BY

JACK FU

IMPORTANT

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any particular purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc., arising out of the use of the information contained herein exceed the fee established for providing such information.

IP7_035315

East Side 476



Praxair
5700 South Alameda Street
Los Angeles, CA 90058
Telephone: (323) 585-2154
Facsimile: (714) 542-6689

CERTIFICATE OF ANALYSIS

CUSTOMER IPSC

DATE 01/26/04

P.O NUMBER

REF. NUMBER 61290100

REQUESTED COMPOSITION

GAS

CONCENTRATION

OXYGEN

8 %

NITROGEN

BALANCE

ANALYTICAL ACCURACY

±1 %

ANALYTICAL METHOD

INSTRUMENT

Siemens Oxymat 5E S/N A12-839

ANALYTICAL PRINCIPLE

Paramagnetic

Values not valid below 150 psig.

This certified mix is N.T.S.T. traceable to SRM#2659a for oxygen.

THIS CYLINDER NO.	CC 110521
CYLINDER PRESSURE	2000 PSIG
EXPIRATION DATE	01/23/07
CLASSIFICATION	PRIMARY MASTER
BATCH NUMBER	N/A
LOT NUMBER	109402009
PART NUMBER	EV NIOX8N-AS
CYLINDER SIZE	AS CGA 590 144 CFT

CERTIFIED CONCENTRATION

OXYGEN	8.01 %
NITROGEN	BALANCE
ANALYTICAL ACCURACY	±1 %

ANALYZED BY

JACK FU

CERTIFIED BY

TY TRIPLETT

IMPORTANT

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any particular purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc., arising out of the use of the information contained herein exceed the fee established for providing such information.

IP7_035316